



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

SPL, Inc.
8820 Interchange Drive
Houston, Tx 77054

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

TESTING

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 08 June 2024

Certificate Number: AT-2877



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

SPL, Inc.
8850 Interchange Drive
Houston, Tx 77054
Brian Eichelberger
Brian.Eichelberger@spl-inc.com
713 660 0901

TESTING

Valid to: **June 8, 2024**

Certificate Number: **AT-2877**

Chemistry


Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Chemical Composition	GPA 2261, GPA 2286	Natural Gas	Gas Chromatography
Chemical Composition	GPA 2177, GPA 2186	Natural Gas Liquid	Gas Chromatography
Chemical Composition	GPA 2103	Natural Gas Condensate	Gas Chromatography
Chemical Composition	ASTM D6667	Natural Gas Liquid	UV Fluorescence
Gases Dissolved in Insulating Liquids	ASTM D3612	Insulating Liquids	Gas Chromatograph
Water Content by Karl Fischer	ASTM D1533	Petroleum Products	Coulometric Titration
Neutralization Number- Acid and Base	ASTM D974	Petroleum Products	Titration
2,6- <i>di-tert</i> -Butyl- <i>p</i> -Cresol and 2,6- <i>di-tert</i> -Butyl Phenol in Insulating Liquids	ASTM D2668	Insulating Liquids	Fourier-transform infrared spectrophotometer
Furanic Coumpounds	ASTM D5837	Insulating Liquids	High-Performance Liquid Chromatography
Interfacial Tension	ASTM D971	Petroleum Products	Tensiometer
ASTM Color	ASTM D1500	Petroleum Products	Colorimeter

Chemistry

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Breakdown Voltage	ASTM D1816	Petroleum Products	Test Cell with VDE electrodes
Density, Relative Density, API Gravity	ASTM D4052	Petroleum Products	Densitometer
Power Factor	ASTM D924	Petroleum Products	Electrical Test Cell

Note:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AT-2877.



Jason Stine, Vice President

